CERTIFICATE OF CALIBRATION

Issued By Transmille Ltd.
Date of Issue EXAMPLE

Certificate Number
EXAMPLE

0324

Approved Signatory

Transmille Ltd.
Unit 4, Select Business Centre
Lodge Road
Staplehurst, Kent. TN12 0QW.
TEL 01580 890700 FAX 01580 890711

www.transmille.com

Customer : EXAMPLE

Date Received : EXAMPLE
Customer Order Number : EXAMPLE

Instrument :

System ID :	EXAMPLE
Description :	Precision Zener Reference
Manufacturer :	Transmille
Model Number :	3000ZR
Serial Number :	EXAMPLE
Procedure Number & Version :	6408 : 1.20/N

Environmental Conditions

Temperature :	20°C ± 1°C	Mains Voltage :	230V ± 12V
Relative Humidity :	40%RH ± 20%RH	Mains Frequency :	50Hz ± 1Hz

Comments

Instrument was placed in lab and allowed to stablise before calibration.

Calibration Information

The instrument was calibrated against laboratory standards whose values are traceable to recognised National Standards. The uncertainty limits quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing a level of confidence of approximatley 95%. The uncertainty evaluation has been carried out in accordance with UKAS requirements.

Calibrated By : EXAMPLE **Date of Calibration : EXAMPLE**

This certificate is issued in accordance with the laboratory accreditation requirements of the United Kingdom Accreditation Service. It provides traceability of measurement to the SI system of units and/or to units of measurement realised at the National Physical Laboratory or other recognised national metrology institutes. This certificate may not be reproduced other than in full, except with the prior written approval of the issuing laboratory.

CERTIFICATE OF CALIBRATION

UKAS Accredited Calibration Laboratory No. **0324**

Certificate Number
EXAMPLE

AFTER ADJUSTMENT RESULTS

Test Title	Applied Value	Reading	Uncertainty
The recorded values are the average of 4 samples			
ZENER BANK 1			
1.018V Reference	1.0180000V	1.0180046V	±1.7uV
10V Reference	10.0000000V	10.0000021V	±10uV
ZENER BANK 2			
1.018V Reference	1.0180000V	1.0180047V	±1.7uV
10V Reference	10.0000000V	9.9999891V	±10uV
ZENER BANK 3			
1.018V Reference	1.0180000V	1.0180053V	±1.7uV
10V Reference	10.0000000V	9.9999994V	±10uV
ZENER BANK 4			
1.018V Reference	1.0180000V	1.0180041V	±1.7uV
10V Reference	10.0000000V	9.9999999V	±10uV

END OF TEST DATA